



Complaint - Revision - Post Market - Feedback Reporting

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs. Contact QA@si-bone.com if you have any questions.

Select the type of problem you are reporting

Problem with individual device or individual patient

General Feedback about SI-BONE Product

Contact Info

Your name Jason Santamore
Your email jason.santamore@si-bone.com
Your phone number (518) 796-1619
Surgeon's name Craig Bartlett

Contact #1: How did you attempt to contact the surgeon? Text

Was this attempt successful? Yes

Comments about attempts to contact surgeon (optional)

He did the revision surgery that was performed by another surgeon a year earlier

Complaint Overview

Date you first heard of problem with SI-BONE product. Friday, July 1, 2022

Which device(s) was/were affected? iFuse-TORQ

Part number(s), if available

N/a

Lot number(s), if available

N/a

Product Complaint or Adverse Event?

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs. Contact QA@si-bone.com if you have any questions.

What is the nature of the problem you are reporting?

Product complaint that caused or may have caused an adverse event or revision procedure

Product Complaint Without Adverse Event

Examples include:

- Damaged instrument
- Broken pin
- Bent pin
- Pin advancement but no patient injury
- Problem with packaging

If a damaged part injured a patient, go back and click on "Product complaint that caused or may have caused an adverse event."

Select Adverse Event Type

Select from the choices below. Your choice guides which section to complete.

If the first 3 do not apply, select "Other".

If a product problem occurred that did NOT cause a patient problem, go back and complete the "Product Complaint Without Adverse Event" section.

What problem did patient have?

Implant malposition causing nerve irritation

Symptomatic Implant Malposition Form

On which side(s) were SI-BONE implants placed during initial surgery?

Right

Which operated side showed postoperative symptoms (i.e., symptoms related to implant malposition)?

Right

Did any SI-BONE staff attend initial surgery?

Yes

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

See summary of IFU steps. Did surgeon complete all steps as shown above?

Yes, all steps were completed accurately

If one or more steps was inaccurately done, describe:

N/a

Were any SI-BONE implants removed or adjusted as a result of this problem?

Yes

Did the patient have SI joint pain relief following the initial procedure? How long was the pain relief ?

Not sure

Date of revision surgery

Thursday, July 21, 2022

Diagnosis / Reason for the iFuse revision surgery?

Implant was on a nerve

Details of iFuse revision surgery (Treatment):

Case went great. Torq removed safely and another was placed between S1 and S2

For above question, be specific and include all steps of the revision surgery and any complications.

- Was the implant adjusted, removed or added?
- Were the explant holes filled with bone graft?
- Were osteotomes and the/or the removal adapter used?
- Was any non-iFuse hardware installed?

Was the iFuse Removal System Instrument Set (P/N 400132) and chisels used to explant the iFuse Implant(s)?

No

Patient's outcome following the iFuse revision surgery: Were the patient's pain complaints resolved?

Not sure

Pain Did Not Improve or Recurred

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Surgical Wound Problem

Other Problem

Use this section ONLY if the patient problem is **NOT**:

- Implant malposition causing nerve irritation
- Pain did not improve or recurred
- Surgical wound problem

Potential Warranty Case

Did patient's situation meet all of the following?

Did patient undergo a revision procedure that met all of the following:

1. *The initial procedure involved placement of 3 SI-BONE implants on the affected side*
2. *The revision procedure:*
 - *Was completed within 1 year of the initial procedure*
 - *Was completed at the same hospital or within the same hospital system as the initial procedure*
 - *Was completed during a separate hospitalization from the initial procedure*
3. *The patient followed all physician instructions after the initial procedure*

Did patient's situation meet warranty criteria shown above?

No, not all of the above apply