



Complaint - Revision - Post Market - Feedback Reporting

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs. Contact QA@si-bone.com if you have any questions.

Select the type of problem you are reporting

Problem with individual device or individual patient

General Feedback about SI-BONE Product

Contact Info

Your name John Parmenter III

Your email john.parmenter@si-bone.com

Your phone number (631) 834-6522

Surgeon's name David Helfet

Contact #1: How did you attempt to contact the surgeon?

Phone

Was this attempt successful?

Yes

Complaint Overview

Date you first heard of problem with SI-BONE product. Friday, June 10, 2022

If problem occurred during placement of an SI-BONE device, enter date placed (initial procedure), not date of revision surgery. Saturday, October 30, 2021

Which device(s) was/were affected?

iFuse-TORQ

Part number(s), if available

ifuse torq 10.0x70 fully threaded
lot# 9046302

Lot number(s), if available

9046302

Product Complaint or Adverse Event?

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs. Contact QA@si-bone.com if you have any questions.

What is the nature of the problem you are reporting?

Product complaint that caused or may have caused an adverse event or revision procedure

Product Complaint Without Adverse Event

Examples include:

- Damaged instrument
- Broken pin
- Bent pin
- Pin advancement but no patient injury
- Problem with packaging

If a damaged part injured a patient, go back and click on "Product complaint that caused or may have caused an adverse event."

Select Adverse Event Type

Select from the choices below. Your choice guides which section to complete.

If the first 3 do not apply, select "Other".

If a product problem occurred that did NOT cause a patient problem, go back and complete the "Product Complaint Without Adverse Event" section.

What problem did patient have?

Implant malposition causing nerve irritation

Symptomatic Implant Malposition Form

On which side(s) were SI-BONE implants placed during initial surgery?

Right

Which operated side showed postoperative symptoms (i.e., symptoms related to implant malposition)?

Right

Did any SI-BONE staff attend initial surgery?

Yes

Name of SI-BONE staff in attendance at initial surgery Gavin Cronin

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
Use of length gauge for implant length selection

- Were standard wound closure techniques used?

See summary of IFU steps. Did surgeon complete all steps as shown above?

Yes, all steps were completed accurately

If one or more steps was inaccurately done, describe:

n/a

Were any SI-BONE implants removed or adjusted as a result of this problem?

Yes

Did the patient have SI joint pain relief following the initial procedure? How long was the pain relief ?

no pain relief

Date of revision surgery

Wednesday, June 22, 2022

Diagnosis / Reason for the iFuse revision surgery?

patient had radiating pelvic pain anterior and posterior

Details of iFuse revision surgery (Treatment):

This was a failed attempt at removing the one malpositioned implant. Surgeon tried backing out the screw with various t handles and trephine for the torq implant.

For above question, be specific and include all steps of the revision surgery and any complications.

- Was the implant adjusted, removed or added?
- Were the explant holes filled with bone graft?
- Were osteotomes and the/or the removal adapter used?
- Was any non-iFuse hardware installed?

Was the iFuse Removal System Instrument Set (P/N 400132) and chisels used to explant the iFuse Implant(s)?

Yes

Patient's outcome following the iFuse revision surgery: Were the patient's pain complaints resolved?

unfortunately this first attempt was a failure in removal. There was a second attempt that was successful.

Pain Did Not Improve or Recurred

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Surgical Wound Problem

Other Problem

Use this section ONLY if the patient problem is **NOT**:

- Implant malposition causing nerve irritation
- Pain did not improve or recurred
- Surgical wound problem

Potential Warranty Case

Did patient's situation meet all of the following?

Did patient undergo a revision procedure that met all of the following:

1. *The initial procedure involved placement of 3 SI-BONE implants on the affected side*
2. *The revision procedure:*
 - *Was completed within 1 year of the initial procedure*
 - *Was completed at the same hospital or within the same hospital system as the initial procedure*
 - *Was completed during a separate hospitalization from the initial procedure*
3. *The patient followed all physician instructions after the initial procedure*

Did patient's situation meet warranty criteria shown above?

No, not all of the above apply

Comment on warranty

patient had the procedure done at a separate hospital near by 8 months after the index procedure.