# **SI-BONE** Complaint - Revision - Post Market - Feedback Reporting

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs. Contact <u>QA@si-bone.com</u> if you have any questions.

# Select the type of problem you are reporting

Problem with individual device or individual patient

# **General Feedback about SI-BONE Product**

This is only for non complaint information. Examples: surgeon suggestions, meeting discussions, etc. For complaints, go back and select "Problem..."

## **Contact Info**

Use this to record your attempts to contact and gather information from the surgeon.

Your name	Morgan Menard-Pailler	
Your email	mmenard-pailler@si-bone.com	
Your phone number	(063) 071-0203	
Surgeon's name	Xavier Loniewski	
Contact #1: How did you attempt to contact the surgeon?	Phone	
Was this attempt successful?	No	
Contact #2: How did you attempt to contact the surgeon?	Email	
Was this attempt successful?	Yes	

#### Comments about attempts to contact surgeon (optional)

He said that the patient want to remove the implant because a persistant radiculalgy pain

## **Complaint Overview**

Enter information about surgery that resulted in a problem with any SI-BONE implant, instrument or accessory (e.g., drill bit, pin).

Date you first heard of problem withWednesday, December 14, 2022SI-BONE product.



If problem occurred during placement Wednesday, October 20, 2021 of an SI-BONE device, enter date placed (initial procedure), not date of revision surgery.

Which device(s) was/were affected?

iFuse

**Part number(s), if available** 7050-100 7040-100 x2

#### Lot number(s), if available

N/A

## **Product Complaint or Adverse Event?**

Decide what type of report you are submitting.

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs. Contact <u>QA@si-bone.com</u> if you have any questions.

# What is the nature of the problem you are reporting?

Product complaint that caused or may have caused an adverse event or revision procedure

# **Product Complaint Without Adverse Event**

Use this form to report an SI-BONE product problem that was NOT associated with a patient adverse event.

Examples include:

- Damaged instrument
- Broken pin
- Bent pin
- Pin advancement but no patient injury
- Problem with packaging

If a damaged part injured a patient, go back and click on "Product complaint that caused or may have caused an adverse event."

#### Date part problem noticed

Wednesday, December 14, 2022

# Select Adverse Event Type

Select from the choices below. Your choice guides which section to complete. If the first 3 do not apply, select "Other". If a product problem occurred that did NOT cause a patient problem, go back and complete the "Product Complaint Without Adverse Event" section.

#### What problem did patient have?

Implant malposition causing nerve irritation

# **Symptomatic Implant Malposition Form**

Use this form if patient experienced symptomatic implant malposition resulting in nerve irritation.



On which side(s) were SI-BONE implants placed during initial surgery?	Left	
Which operated side showed postoperative symptoms (i.e., symptoms related to implant malposition)?	Left	
Did any SI-BONE staff attend initial surgery?	Yes	

# Name of SI-BONE staff in attendance Olivier Terrasson at initial surgery

(Information should be from SI-BONE staff who attended initial surgery) The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment Use of length gauge for implant length selection
- Were standard wound closure techniques used?

# See summary of IFU steps. Did surgeon complete all steps as shown above?

Yes, all steps were completed accurately

#### If one or more steps was inaccurately done, describe:

N/A

Were any SI-BONE implants removed or adjusted as a result of this problem?

Did the patient have SI joint pain relief after few weeks following the initial procedure? How long was the pain relief ?

For above question, be specific and include all steps of the revision surgery and any complications.

No

- Was the implant adjusted, removed or added?
- Were the explant holes filled with bone graft?
- Were osteotomes and the/or the removal adapter used?
- Was any non-iFuse hardware installed?

## **Pain Did Not Improve or Recurred**

Use this form if pain did not improve OR pain improved but then recurred

(Information should be from SI-BONE staff who attended initial surgery) The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use



- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment Use of length gauge for implant length selection
- Were standard wound closure techniques used?

# **Surgical Wound Problem**

Use this form if patient developed postoperative surgical wound problem (infection, dehiscence).

# **Other Problem**

Use this section ONLY if the patient problem is NOT:

- Implant malposition causing nerve irritation
- Pain did not improve or recurred
- Surgical wound problem

## **Potential Warranty Case**

Please help evaluate whether this case could qualify for SI-BONE's warranty policy.

Did patient's situation meet all of the following?

- Did patient undergo a revision procedure that met all of the following:
  - 1. The initial procedure involved placement of 3 SI-BONE implants on the affected side
  - 2. The revision procedure:
    - Was completed within 1 year of the initial procedure
    - Was completed at the same hospital or within the same hospital system as the initial procedure
    - Was completed during a separate hospitalization from the initial procedure
  - 3. The patient followed all physician instructions after the initial procedure

# Did patient's situation meet warranty criteria shown above?

No, not all of the above apply

#### **Comment on warranty**

The revision procedure was after 1 year after the initial procedure in a different hospital but by the same surgeon

