

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs.

Contact **QA@si-bone.com** if you have any questions.

Contact Info

Use this to record your attempts to contact and gather information from the customer

Your name Megan Hinkle

Customer name Kraig Kristof

How did you learn about this issue? (select all that apply)?

From the HCP or associated staff

Please provide any relevant details about your communication. Full complaint description will be captured on the following page:

Patient had original SIJF 10/7/2020. Patient complained of pain with positive fortin finger. Injections provided relief.

Complaint Overview

Enter information about surgery that resulted in a problem with any SI-BONE implant, instrument or accessory (e.g., drill bit, pin).

Date you first heard of problem with SI-BONE product.

Friday, January 26, 2024

Date of original surgery (if revision is being reported) or alleged product failure

Wednesday, October 7, 2020

Indicate affected device(s) (choose all that apply)

iFuse-3D

Part number(s) (please list the number of each part involved)(required)

7045M-90

Lot number(s)

2731781

Product Complaint or Adverse Event?

Decide what type of report you are submitting. Contact QA@si-bone.com if you have any questions.

YES, potential or actual (Ex: required revision, patient adverse event)

Product Complaint Without Patient Problem

Use this form to report an SI-BONE product problem that was NOT associated with a patient adverse event.

Examples include:

- Damaged instrument/implant
- Broken/ bent/ cut pin
- · Pin advancement but no patient injury
- · Packaging issue

If patient injury occured, go back and click YES to report patient problem.

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
 Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Select Adverse Event Type

What problem did patient have?

Continued, recurrent, or new pain

Implant Malposition Form

Use this form if patient an implant malposition was detected.

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
 Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Continued, recurrent, or new pain

Use this form if pain did not improve, pain improved but then returned, or new onset pain

Best description of time course of pain recurrence:

Pain got better but then recurred

How long did the patient experience pain relief?

More than 12 months

Were any additional causes of pain discovered during workup?

I don't know

If CT was performed, please email scan to QA@si-bone.com. CT results show:

CT was done, but doctor refuses to comment on results

Was initial surgery attended by SI-BONE staff member?

Yes

Name of SI-BONE staff member attending initial surgery

Mike Schwamel

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
 Use of length gauge for implant length selection
- Were standard wound closure techniques used?

See summary of IFU steps. Did surgeon complete all steps as shown above?

No, one or more steps was inaccurately done. (Ignore the fact that patient had malpositioned implant. Focus on whether procedure was correctly executed.)

Please describe any steps inaccurately performed, or other details of the case

Only 2 implants were used in the case.

Did patient have revision surgery as a result of this problem?

Yes

Surgical Wound Problem

Use this form if patient developed postoperative surgical wound problem (infection, dehiscence).

Other Problem

Use this section ONLY if the patient problem is NOT:

- Implant malposition causing nerve irritation
- · Pain did not improve or recurred
- Surgical wound problem

Revision Procedure

Complete this form as thoroughly as possible for a complaint involving a revision surgery.

Please indicate date of revision procedure

Monday, February 5, 2024

Reason for revision (e.g. nerve impingement, loosening, etc.). Please be as specific as possible:

Pseudoarthosis

Which step(s) were performed during the revision? Choose all that apply:

iFuse implant was removed

Additional iFuse implant was placed

Please further describe the revision procedure (any issues with instrumentation or medical issues?. Be as specific as possible. Failure to provide details will result in continued follow up with you:

Inferrior implant was removed using the removal system and replaced with a 10.75 mm implant. A 3rd 7.0 iFuse 3D implant was also placed.

To your knowledge, was the patient's issue resolved after surgery?

Unknown

You may be contacted for further information if your submission is lacking critical details. We appreciate your thoroughness.