



Complaint Reporting

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs.

Contact QA@si-bone.com if you have any questions.

Contact Info

Use this to record your attempts to contact and gather information from the customer

Your name Ryan Perestock

Customer name Nestor Tomycz

How did you learn about this issue? (select all that apply)? From the HCP or associated staff

Please provide any relevant details about your communication. Full complaint description will be captured on the following page:

Dr. Tomycz informed me about the removal case the week prior to surgery.

Complaint Overview

Enter information about surgery that resulted in a problem with any SI-BONE implant, instrument or accessory (e.g., drill bit, pin).

Date you first heard of problem with SI-BONE product. Friday, July 19, 2024

Date of original surgery (if revision is being reported) or alleged product failure Friday, November 17, 2023

Indicate affected device(s) (choose all that apply) iFuse-3D

Part number(s) (please list the number of each part involved)(required)

- 7055M-90
- 7040M-90
- 7035M-90

Lot number(s)

- 7055M-90 - 9084541
- 7040M-90 - 9085931
- 7035M-90 - 9084371

Product Complaint or Adverse Event?

Decide what type of report you are submitting. Contact QA@si-bone.com if you have any questions.

Did the product complaint result in a patient problem?

YES, potential or actual (Ex: required revision, patient adverse event)

Product Complaint Without Patient Problem

Use this form to report an SI-BONE product problem that was NOT associated with a patient adverse event.

Examples include:

- Damaged instrument/implant
- Broken/ bent/ cut pin
- Pin advancement but no patient injury
- Packaging issue

If patient injury occurred, go back and click YES to report patient problem.

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Select Adverse Event Type

What problem did patient have?

Implant malposition (e.g. causing nerve irritation)

Implant Malposition Form

Use this form if patient an implant malposition was detected.

On which side(s) were SI-BONE implants placed during initial surgery?

Left

Which side shows implant malposition?

Left

Did any SI-BONE staff attend initial surgery?

Yes

Name of SI-BONE staff in attendance at initial surgery

Ryan Perestock

Imaging type used during initial surgery

C-arm/fluoro only

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

See summary of IFU steps. Did surgeon complete all steps as shown above (regardless of post-op symptoms)?

Yes, all steps were completed accurately

Please describe procedure steps not done properly or other pertinent information

Primary case on 11/17/23 resulted in removing the 7055M-90 implant on 11/19/23. Dr. Tomycz informed me the patients issue was resolved but a few months later starting experiencing foot pain in certain positions standing and sitting.

Did patient have revision surgery?

Yes

Continued, recurrent, or new pain

Use this form if pain did not improve, pain improved but then returned, or new onset pain

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Surgical Wound Problem

Use this form if patient developed postoperative surgical wound problem (infection, dehiscence).

Other Problem

Use this section ONLY if the patient problem is **NOT**:

- Implant malposition causing nerve irritation
- Pain did not improve or recurred
- Surgical wound problem

Revision Procedure

Complete this form as thoroughly as possible for a complaint involving a revision surgery.

Please indicate date of revision procedure

Thursday, July 25, 2024

Reason for revision (e.g. nerve impingement, loosening, etc.). Please be as specific as possible:

The patient's SIJ pain was resolved but started experiencing foot pain/nerve pain in certain positions. Dr. Tomycz concluded the best option was to remove the remaining 2 iFuse 3D implants.

Which step(s) were performed during the revision? Choose all that apply:

iFuse implant was removed

Please further describe the revision procedure (any issues with instrumentation or medical issues?). Be as specific as possible. Failure to provide details will result in continued follow up with you:

Standard removal surgical technique steps were followed to remove the 7040M-90 and 7035M-90 implants. The removal chisel guide was used to chisel all 3 sides of each implant and they were successfully removed.

To your knowledge, was the patient's issue resolved after surgery?

Unknown

You may be contacted for further information if your submission is lacking critical details. We appreciate your thoroughness.