



Complaint Reporting

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs.

Contact QA@si-bone.com if you have any questions.

Contact Info

Use this to record your attempts to contact and gather information from the customer

Your name Jack Wohlabaugh

Customer name Andrew Blankenship

How did you learn about this issue? (select all that apply)? From the HCP or associated staff

Please provide any relevant details about your communication. Full complaint description will be captured on the following page:

Dr. Blankenship reached out to us a couple days after the surgery telling us the patient was experiencing a lot of pain and went back to the emergency room twice. After further evaluation Dr. Blankenship told us there was an implant violating the S1 foramen

Complaint Overview

Enter information about surgery that resulted in a problem with any SI-BONE implant, instrument or accessory (e.g., drill bit, pin).

Date you first heard of problem with SI-BONE product. Wednesday, November 27, 2024

Date of original surgery (if revision is being reported) or alleged product failure Wednesday, November 13, 2024

Indicate affected device(s) (choose all that apply) iFuse-TORQ

Part number(s) (please list the number of each part involved)(required)
N/a

Lot number(s)
9084231

Product Complaint or Adverse Event?

Decide what type of report you are submitting. Contact QA@si-bone.com if you have any questions.

Did the product complaint result in a patient problem?

YES, potential or actual (Ex: required revision, patient adverse event)

Product Complaint Without Patient Problem

Use this form to report an SI-BONE product problem that was NOT associated with a patient adverse event.

Examples include:

- Damaged instrument/implant
- Broken/ bent/ cut pin
- Pin advancement but no patient injury
- Packaging issue

If patient injury occurred, go back and click YES to report patient problem.

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Select Adverse Event Type

What problem did patient have?

Continued, recurrent, or new pain

Implant Malposition Form

Use this form if patient an implant malposition was detected.

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Continued, recurrent, or new pain

Use this form if pain did not improve, pain improved but then returned, or new onset pain

Best description of time course of pain recurrence:

New pain, different from pre-op

How long did the patient experience pain relief?

Less than 6 months

Were any additional causes of pain discovered during workup?

I don't know

Describe discovered or suspected other causes of pain

Torq implant was violating the s1 foramen

If CT was performed, please email scan to QA@si-bone.com. CT results show:

No CT was done

Additional CT results / details

Ct is available and being sent

Was initial surgery attended by SI-BONE staff member?

Yes

Name of SI-BONE staff member attending initial surgery

Jack Wohlabaugh

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

See summary of IFU steps. Did surgeon complete all steps as shown above?

Yes, all steps were completed accurately

Please describe any steps inaccurately performed, or other details of the case

None

Did patient have revision surgery as a result of this problem?

Yes

Surgical Wound Problem

Use this form if patient developed postoperative surgical wound problem (infection, dehiscence).

Other Problem

Use this section ONLY if the patient problem is **NOT**:

- Implant malposition causing nerve irritation
- Pain did not improve or recurred
- Surgical wound problem

Revision Procedure

Complete this form as thoroughly as possible for a complaint involving a revision surgery.

Please indicate date of revision procedure

Tuesday, December 3, 2024

Reason for revision (e.g. nerve impingement, loosening, etc.). Please be as specific as possible:

Implant in s1 foramen causing patient a lot of pain and discomfort

Which step(s) were performed during the revision? Choose all that apply:

iFuse implant was removed

Additional iFuse implant was placed

Please further describe the revision procedure (any issues with instrumentation or medical issues?). Be as specific as possible. Failure to provide details will result in continued follow up with you:

We took out the original implant which was an 11.5x55. It was slightly too long and placed where it was just violating the s1 foramen. We took it out and replaced it with a 13.5x45

To your knowledge, was the patient's issue resolved after surgery?

Yes

You may be contacted for further information if your submission is lacking critical details. We appreciate your thoroughness.