



Complaint Reporting

Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs.

Contact QA@si-bone.com if you have any questions.

Contact Info

Use this to record your attempts to contact and gather information from the customer

Your name Amanda Adams

Customer name Paul Krafft

**How did you learn about this issue?
(select all that apply)?** From the HCP or associated staff

Please provide any relevant details about your communication. Full complaint description will be captured on the following page:

Halifax Health reached out to me the morning of 12/26/2024 and informed me of a Granite revision case that same day.

Complaint Overview

Enter information about surgery that resulted in a problem with any SI-BONE implant, instrument or accessory (e.g., drill bit, pin).

Date you first heard of problem with SI-BONE product. Thursday, December 26, 2024

Date of original surgery (if revision is being reported) or alleged product failure Wednesday, October 25, 2023

Indicate affected device(s) (choose all that apply) iFuse Bedrock Granite

Part number(s) (please list the number of each part involved)(required)
105100BG and 105100BG

Lot number(s)
SN 29961022302-196 and Lot 29960532303

Product Complaint or Adverse Event?

Decide what type of report you are submitting. Contact QA@si-bone.com if you have any questions.

Did the product complaint result in a patient problem?

YES, potential or actual (Ex: required revision, patient adverse event)

Product Complaint Without Patient Problem

Use this form to report an SI-BONE product problem that was NOT associated with a patient adverse event.

Examples include:

- Damaged instrument/implant
- Broken/ bent/ cut pin
- Pin advancement but no patient injury
- Packaging issue

If patient injury occurred, go back and click YES to report patient problem.

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Select Adverse Event Type

What problem did patient have?

Other problem

Implant Malposition Form

Use this form if patient an implant malposition was detected.

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
- Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Continued, recurrent, or new pain

Use this form if pain did not improve, pain improved but then returned, or new onset pain

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment
 - Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Surgical Wound Problem

Use this form if patient developed postoperative surgical wound problem (infection, dehiscence).

Other Problem

Use this section ONLY if the patient problem is **NOT**:

- Implant malposition causing nerve irritation
- Pain did not improve or recurred
- Surgical wound problem

Describe problem in detail

The patient picked up a heavy bag of concrete and felt a pop in his back and new pain. CT scan showed both set screws popped of the granite implants

Did patient undergo revision surgery to address this problem?

Yes

Revision Procedure

Complete this form as thoroughly as possible for a complaint involving a revision surgery.

Please indicate date of revision procedure

Thursday, December 26, 2024

Reason for revision (e.g. nerve impingement, loosening, etc.). Please be as specific as possible:

New pain and both set screws popped off granite implants

Which step(s) were performed during the revision? Choose all that apply:

Additional iFuse implant was placed

New setscrews were placed on existing granite implants. Added Granite at S1 bilateral and one iliac granite implant

Please further describe the revision procedure (any issues with instrumentation or medical issues?. Be as specific as possible. Failure to provide details will result in continued follow up with you:

Spinewave S1 implants failed and were replaced by granite 9.5's. Added 1 granite iliac implant with another rod. Both original rods were also replaced.

To your knowledge, was the patient's issue resolved after surgery?

Unknown

You may be contacted for further information if your submission is lacking critical details. We appreciate your thoroughness.