Monday, February 17, 2025



Use this form to submit information about a complaint related to an SI-BONE product to Quality Affairs.

Contact <u>QA@si-bone.com</u> if you have any questions.

Contact Info

Use this to record your attempts to contact and gather information from the customer

Your name

Bethann Barry

Customer name

Kevin Yoo

How did you learn about this issue? (select all that apply)?

I observed the issue

Please provide any relevant details about your communication. Full complaint description will be captured on the following page:

Set Screws pooped off

Complaint Overview

Enter information about surgery that resulted in a problem with any SI-BONE implant, instrument or accessory (e.g., drill bit, pin).

Date you first heard of problem with SI-BONE product.	Sunday, February 9, 2025
Date of original surgery (if revision is being reported) or alleged product failure	Sunday, September 17, 2023
Indicate affected device(s) (choose all that apply)	iFuse Bedrock Granite

Part number(s) (please list the number of each part involved)(required)

105100BG - 29961282305-068 105100BG - 29961282305-009

Lot number(s)

105100BG - 29961282305-068 105100BG - 29961282305-009

Product Complaint or Adverse Event?

Decide what type of report you are submitting. Contact QA@si-bone.com if you have any questions.

Product Complaint Without Patient Problem

Use this form to report an SI-BONE product problem that was NOT associated with a patient adverse event.

Examples include:

- Damaged instrument/implant
- Broken/ bent/ cut pin
- Pin advancement but no patient injury
- Packaging issue

If patient injury occured, go back and click YES to report patient problem.

When was problem detected?

set screws popped off after revision surgery of initial hardware

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Please describe the details of the event as fully as possible

Patient had an extension down to the pelvis o 9/13/23

On 1/31 Patient was brought back due to hardware failures and all screws were changed out. Granite stayed in.

Granite set screws popped off and was revised adding more instrumentation on 2/9. One Granite was able to come out and replaced with an 11.5

Select Adverse Event Type

Implant Malposition Form

Use this form if patient an implant malposition was detected.

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Continued, recurrent, or new pain

Use this form if pain did not improve, pain improved but then returned, or new onset pain

(Information should be from SI-BONE staff who attended initial surgery)

The surgeon completed all steps in the IFU, including:

- Patient setup (table, spine neutral position, prone or supine position)
- Inspection of instruments and implants prior to use
- Proper views (lateral, inlet and outlet) obtained during use of instruments and placement of implants that clearly showed recommended anatomic landmarks
- Appropriate implant starting position
- (If performed) proper use of neuromonitoring guide pin and equipment Use of length gauge for implant length selection
- Were standard wound closure techniques used?

Surgical Wound Problem

Use this form if patient developed postoperative surgical wound problem (infection, dehiscence).

Other Problem

Use this section ONLY if the patient problem is **NOT**:

- Implant malposition causing nerve irritation
- Pain did not improve or recurred
- Surgical wound problem

Revision Procedure

Complete this form as thoroughly as possible for a complaint involving a revision surgery.

You may be contacted for further information if your submission is lacking critical details. We appreciate your thoroughness.